

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020823\
 Data File : VN076588.D
 Acq On : 08 Feb 2023 18:34
 Operator : JC\MD
 Sample : 01418-10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JPP-20-020223

Quant Time: Feb 09 00:53:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020723W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 08 09:04:43 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.233	168	652334	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.104	114	1116436	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.868	117	987667	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.792	152	382387	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.586	65	375329	48.747	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.500%
35) Dibromofluoromethane	8.175	113	330472	47.618	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.240%
50) Toluene-d8	10.569	98	1297197	49.684	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.360%
62) 4-Bromofluorobenzene	12.851	95	416653	45.298	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.600%
Target Compounds						
					Qvalue	
16) Acetone	4.451	43	38230	15.060	ug/l	99
20) Methylene Chloride	5.298	84	28260	3.611	ug/l	98
37) Ethyl Acetate	7.575	43	58967	7.943	ug/l	96
43) Isopropyl Acetate	8.698	43	326159	26.403	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020823\
Data File : VN076588.D
Acq On : 08 Feb 2023 18:34
Operator : JC\MD
Sample : 01418-10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JPP-20-020223

Quant Time: Feb 09 00:53:10 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020723W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 08 09:04:43 2023
Response via : Initial Calibration

