



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020823\
Data File : VN076566.D
Acq On : 08 Feb 2023 09:06
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020723W.M
Title : SW846 8260
Last Update : Wed Feb 08 09:04:43 2023



AutoFind: Scans 1835, 1836, 1837; Background Corrected with Scan 1826

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.4	24467	PASS
75	95	30	60	48.7	77299	PASS
95	95	100	100	100.0	158741	PASS
96	95	5	9	6.4	10191	PASS
173	174	0.00	2	0.8	1012	PASS
174	95	50	100	76.4	121275	PASS
175	174	5	9	6.8	8223	PASS
176	174	95	101	97.0	117653	PASS
177	176	5	9	6.6	7744	PASS

Average of 12.847 to 12.859 min.: VN076566.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
36.00	1048	50.10	24467	63.00	5281	76.05	6811
37.05	5745	51.05	8430	64.00	609	77.00	981
38.05	5229	51.90	134	65.05	606	78.05	591
39.05	2320	52.15	180	66.95	523	78.95	476
43.10	127	54.95	457	68.00	15047	79.95	1260
44.00	513	55.95	2088	69.05	15598	80.95	4906
44.95	1159	57.00	3313	70.00	1092	81.95	968
46.00	71	58.00	88	71.95	839	83.10	76
46.95	1377	60.05	1342	73.05	7023	85.90	81
48.00	872	61.05	6793	74.10	25933	87.05	6775
49.05	5671	62.00	7084	75.10	77299	88.00	6336

Average of 12.847 to 12.859 min.: VN076566.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
90.95	513	105.95	764	127.95	453	143.00	1952
92.05	4633	107.00	83	128.85	325	144.80	149
93.05	6824	112.70	63	129.80	341	145.05	238
94.05	18417	113.20	57	130.05	323	145.80	94
95.10	158741	114.90	178	130.80	89	146.00	94
96.05	10191	115.90	695	130.95	183	147.90	433
97.05	293	116.95	960	134.70	63	149.90	61
102.90	62	117.90	575	135.05	207	152.90	159
103.90	682	119.00	882	136.95	189	154.00	68
104.75	179	124.00	53	141.05	1742	154.90	165
105.10	89	127.80	222	141.85	213	155.05	229

Average of 12.847 to 12.859 min.: VN076566.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
156.90	295	177.05	7744				
158.95	252	177.95	159				
160.80	67	281.30	50				
171.00	106						
172.00	396						
172.20	507						
172.80	281						
173.05	1012						
174.00	121275						
175.00	8223						
176.00	117653						