

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020824\
 Data File : VN080969.D
 Acq On : 08 Feb 2024 17:22
 Operator : JC\MD
 Sample : P1372-08 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CG-MW-17-ch7

Quant Time: Feb 09 01:24:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020624W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 07 06:22:23 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	497372	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	933973	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	815897	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	317619	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	344178	54.855	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.700%	
35) Dibromofluoromethane	8.171	113	283623	56.592	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 113.180%	
50) Toluene-d8	10.571	98	985328	54.355	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 108.700%	
62) 4-Bromofluorobenzene	12.853	95	325798	49.473	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 98.940%	
Target Compounds						
					Qvalue	
40) Benzene	8.612	78	889015	39.173	ug/l	95
67) Ethyl Benzene	11.965	91	75549	2.565	ug/l	96
68) m/p-Xylenes	12.082	106	3417	0.305	ug/l	95
69) o-Xylene	12.406	106	4350	0.401	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020824\
Data File : VN080969.D
Acq On : 08 Feb 2024 17:22
Operator : JC\MD
Sample : P1372-08 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CG-MW-17-ch7

Quant Time: Feb 09 01:24:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020624W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 07 06:22:23 2024
Response via : Initial Calibration

