

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020921\
 Data File : VN065773.D
 Acq On : 9 Feb 2021 11:33
 Operator : JC/MD
 Sample : VN0209WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0209WBL01

Quant Time: Feb 09 14:17:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	186304	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.547	114	307945	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	272163	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.312	152	96133	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.984	65	104908	42.80	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	85.60%
35) Dibromofluoromethane	7.547	113	88494	48.49	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.98%
50) Toluene-d8	10.055	98	380408	53.03	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.06%
62) 4-Bromofluorobenzene	12.373	95	120977	46.83	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	93.66%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020921\
Data File : VN065773.D
Acq On : 9 Feb 2021 11:33
Operator : JC/MD
Sample : VN0209WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0209WBL01

Quant Time: Feb 09 14:17:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

