

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020921\
 Data File : VN065776.D
 Acq On : 9 Feb 2021 13:06
 Operator : JC/MD
 Sample : M1318-03 10X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 31408-31410

Quant Time: Feb 09 14:19:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.621	168	169722	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.547	114	280260	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	248499	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.312	152	92041	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.984	65	94459	42.30	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	84.60%
35) Dibromofluoromethane	7.547	113	82299	49.55	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.10%
50) Toluene-d8	10.055	98	348378	53.36	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.72%
62) 4-Bromofluorobenzene	12.370	95	112336	47.79	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.58%
Target Compounds						
					Qvalue	
16) Acetone	3.778	43	12762	14.68	ug/l	100
18) Methyl Acetate	4.283	43	18438	7.57	ug/l	96
40) Benzene	8.000	78	12366	1.44	ug/l	98
52) Toluene	10.119	92	11526	2.26	ug/l	96
68) m/p-Xylenes	11.588	106	6637	1.88	ug/l	98
95) Naphthalene	15.100	128	6529	4.47	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020921\
Data File : VN065776.D
Acq On : 9 Feb 2021 13:06
Operator : JC/MD
Sample : M1318-03 10X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
31408-31410

Quant Time: Feb 09 14:19:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

