

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020921\
 Data File : VN065779.D
 Acq On : 9 Feb 2021 14:45
 Operator : JC/MD
 Sample : PB134480TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB134480TB

Quant Time: Feb 10 00:58:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	172035	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.547	114	282770	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	252471	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.312	152	92667	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	96629	42.69	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	85.38%
35) Dibromofluoromethane	7.547	113	80218	47.87	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.74%
50) Toluene-d8	10.058	98	349917	53.12	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.24%
62) 4-Bromofluorobenzene	12.373	95	114467	48.26	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	96.52%
Target Compounds						
16) Acetone	3.782	43	16666	18.92	ug/l	98
20) Methylene Chloride	4.496	84	9442	3.87	ug/l	95
43) Isopropyl Acetate	8.132	43	12876	2.66	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020921\
Data File : VN065779.D
Acq On : 9 Feb 2021 14:45
Operator : JC/MD
Sample : PB134480TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB134480TB

Quant Time: Feb 10 00:58:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

