

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020921\
 Data File : VN065780.D
 Acq On : 9 Feb 2021 15:10
 Operator : JC/MD
 Sample : PB134530TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB134530TB

Quant Time: Feb 10 00:58:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	172853	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	288298	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	260179	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	95292	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.984	65	98166	43.17	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	86.34%
35) Dibromofluoromethane	7.547	113	82232	48.13	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.26%
50) Toluene-d8	10.058	98	358078	53.32	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.64%
62) 4-Bromofluorobenzene	12.373	95	118496	49.00	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	98.00%
Target Compounds						
16) Acetone	3.779	43	15240	17.21	ug/l	90
20) Methylene Chloride	4.495	84	9530	3.89	ug/l	91
43) Isopropyl Acetate	8.129	43	12734	2.58	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020921\
Data File : VN065780.D
Acq On : 9 Feb 2021 15:10
Operator : JC/MD
Sample : PB134530TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB134530TB

Quant Time: Feb 10 00:58:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

