

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020921\
 Data File : VN065782.D
 Acq On : 9 Feb 2021 15:59
 Operator : JC/MD
 Sample : PB134530ZHE#02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB134530ZHE#02

Quant Time: Feb 10 00:59:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	173556	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.547	114	292479	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	264156	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	96244	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.984	65	99839	43.73	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	87.46%		
35) Dibromofluoromethane	7.547	113	82197	47.42	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	94.84%		
50) Toluene-d8	10.058	98	364576	53.51	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	107.02%		
62) 4-Bromofluorobenzene	12.373	95	119482	48.70	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	97.40%		
Target Compounds						
						Qvalue
16) Acetone	3.782	43	14766	16.61	ug/l	96
18) Methyl Acetate	4.287	43	3022	1.21	ug/l #	76
20) Methylene Chloride	4.502	84	10064	4.09	ug/l	92
43) Isopropyl Acetate	8.132	43	12819	2.56	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020921\
Data File : VN065782.D
Acq On : 9 Feb 2021 15:59
Operator : JC/MD
Sample : PB134530ZHE#02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB134530ZHE#02

Quant Time: Feb 10 00:59:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

