

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020921\
 Data File : VN065786.D
 Acq On : 9 Feb 2021 17:38
 Operator : JC/MD
 Sample : M1331-05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ETGI-351-360

Quant Time: Feb 10 01:00:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	179124	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.547	114	296836	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	270792	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.312	152	96936	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	99270	42.13	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	84.26%
35) Dibromofluoromethane	7.547	113	84924	48.28	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.56%
50) Toluene-d8	10.055	98	368924	53.36	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.72%
62) 4-Bromofluorobenzene	12.373	95	119370	47.94	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.88%
Target Compounds						
16) Acetone	3.782	43	18262	19.91	ug/l	92
18) Methyl Acetate	4.290	43	4119	1.60	ug/l #	67
20) Methylene Chloride	4.496	84	8405	3.31	ug/l	93
43) Isopropyl Acetate	8.129	43	6289	1.24	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020921\
Data File : VN065786.D
Acq On : 9 Feb 2021 17:38
Operator : JC/MD
Sample : M1331-05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ETGI-351-360

Quant Time: Feb 10 01:00:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

