

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020921\
 Data File : VN065787.D
 Acq On : 9 Feb 2021 18:03
 Operator : JC/MD
 Sample : M1339-05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-0-5

Quant Time: Feb 10 01:00:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	178150	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.547	114	293179	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	269271	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	97222	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.984	65	97867	41.76	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	83.52%
35) Dibromofluoromethane	7.550	113	82785	47.65	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.30%
50) Toluene-d8	10.055	98	368125	53.90	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.80%
62) 4-Bromofluorobenzene	12.370	95	119343	48.53	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	97.06%
Target Compounds						
16) Acetone	3.775	43	13163	14.43	ug/l	95
18) Methyl Acetate	4.296	43	3913	1.53	ug/l #	73
20) Methylene Chloride	4.502	84	8818	3.49	ug/l	95
43) Isopropyl Acetate	8.129	43	11759	2.34	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020921\
Data File : VN065787.D
Acq On : 9 Feb 2021 18:03
Operator : JC/MD
Sample : M1339-05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-0-5

Quant Time: Feb 10 01:00:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

