

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020921\
 Data File : VN065788.D
 Acq On : 9 Feb 2021 18:27
 Operator : JC/MD
 Sample : M1339-10
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-5-10

Manual Integrations
 APPROVED

MMDadoda
 2/10/2021 10:08:00 AM

Quant Time: Feb 10 01:00:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	179465	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.547	114	299725	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	271982	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	100904	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	100476	42.56	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	85.12%		
35) Dibromofluoromethane	7.547	113	86218	48.54	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	97.08%		
50) Toluene-d8	10.055	98	371518	53.21	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	106.42%		
62) 4-Bromofluorobenzene	12.373	95	121652	48.39	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	96.78%		
Target Compounds					Qvalue	
16) Acetone	3.779	43	13259	14.43	ug/l	95
18) Methyl Acetate	4.290	43	3844	1.49	ug/l	93
20) Methylene Chloride	4.505	84	9493m	3.73	ug/l	
43) Isopropyl Acetate	8.129	43	12363	2.41	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020921\
Data File : VN065788.D
Acq On : 9 Feb 2021 18:27
Operator : JC/MD
Sample : M1339-10
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-5-10

Manual Integrations
APPROVED

MMDadoda
2/10/2021 10:08:00 AM

Quant Time: Feb 10 01:00:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

