

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021020\
 Data File : VN060041.D
 Acq On : 10 Feb 2020 13:17
 Operator : JC/MD
 Sample : L1438-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 HAM-WC-S-SB-0203

Quant Time: Feb 11 07:35:16 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 18 03:09:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	204983	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	321310	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	299151	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	131520	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	118008	51.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.68%	
35) Dibromofluoromethane	7.57	113	97087	50.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.34%	
50) Toluene-d8	10.08	98	388718	52.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.72%	
62) 4-Bromofluorobenzene	12.40	95	137319	51.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.62%	
Target Compounds						
16) Acetone	3.78	43	41238	44.901	ug/l	99
20) Methylene Chloride	4.52	84	5280	1.303	ug/l	90
25) 2-Butanone	6.81	43	7145	5.543	ug/l #	86
43) Isopropyl Acetate	8.14	43	46028	10.197	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021020\
Data File : VN060041.D
Acq On : 10 Feb 2020 13:17
Operator : JC/MD
Sample : L1438-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
HAM-WC-S-SB-0203

Quant Time: Feb 11 07:35:16 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
Quant Title : SW846 8260
QLast Update : Sat Jan 18 03:09:32 2020
Response via : Initial Calibration

