

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021020\
 Data File : VN060042.D
 Acq On : 10 Feb 2020 13:40
 Operator : JC/MD
 Sample : L1456-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ETGI-337-290

Quant Time: Feb 11 07:36:33 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 18 03:09:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	200088	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	312500	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	289561	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	129816	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	114714	51.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.26%	
35) Dibromofluoromethane	7.57	113	95110	51.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.08%	
50) Toluene-d8	10.08	98	375519	52.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.02%	
62) 4-Bromofluorobenzene	12.40	95	134666	51.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.48%	
Target Compounds						
16) Acetone	3.79	43	12075	6.169	ug/l	96
20) Methylene Chloride	4.53	84	4873	1.160	ug/l	88
43) Isopropyl Acetate	8.14	43	48577	11.065	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021020\
Data File : VN060042.D
Acq On : 10 Feb 2020 13:40
Operator : JC/MD
Sample : L1456-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ETGI-337-290

Quant Time: Feb 11 07:36:33 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
Quant Title : SW846 8260
QLast Update : Sat Jan 18 03:09:32 2020
Response via : Initial Calibration

