

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021020\
 Data File : VN060046.D
 Acq On : 10 Feb 2020 15:11
 Operator : JC/MD
 Sample : PB126661ZHE#13
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB126661ZHE#13

Quant Time: Feb 11 07:44:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 18 03:09:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	196693	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	306846	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	291636	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	135884	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	114041	52.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.42%	
35) Dibromofluoromethane	7.57	113	93460	51.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.16%	
50) Toluene-d8	10.08	98	371762	52.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.88%	
62) 4-Bromofluorobenzene	12.40	95	136405	53.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.74%	
Target Compounds						
20) Methylene Chloride	4.53	84	5304	1.427	ug/l	85
43) Isopropyl Acetate	8.14	43	52420	12.160	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021020\
Data File : VN060046.D
Acq On : 10 Feb 2020 15:11
Operator : JC/MD
Sample : PB126661ZHE#13
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB126661ZHE#13

Quant Time: Feb 11 07:44:41 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
Quant Title : SW846 8260
QLast Update : Sat Jan 18 03:09:32 2020
Response via : Initial Calibration

