

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021020\
 Data File : VN060049.D
 Acq On : 10 Feb 2020 16:19
 Operator : JC/MD
 Sample : PB126661ZHE#16
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB126661ZHE#16

Quant Time: Feb 11 07:48:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 18 03:09:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	205169	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	318237	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	302958	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	139359	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	117807	51.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.40%	
35) Dibromofluoromethane	7.57	113	96293	50.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.48%	
50) Toluene-d8	10.08	98	387363	52.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.36%	
62) 4-Bromofluorobenzene	12.40	95	141105	53.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.46%	
Target Compounds						
20) Methylene Chloride	4.52	84	6796	2.056	ug/l	89
43) Isopropyl Acetate	8.14	43	50798	11.362	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021020\
Data File : VN060049.D
Acq On : 10 Feb 2020 16:19
Operator : JC/MD
Sample : PB126661ZHE#16
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB126661ZHE#16

Quant Time: Feb 11 07:48:27 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
Quant Title : SW846 8260
QLast Update : Sat Jan 18 03:09:32 2020
Response via : Initial Calibration

