

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021020\
 Data File : VN060050.D
 Acq On : 10 Feb 2020 16:42
 Operator : JC/MD
 Sample : PB126661ZHE#17
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB126661ZHE#17

Quant Time: Feb 11 07:51:10 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 18 03:09:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	199106	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	310899	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	287446	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	128410	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	115670	52.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.62%	
35) Dibromofluoromethane	7.57	113	95871	51.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.42%	
50) Toluene-d8	10.08	98	377962	52.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.24%	
62) 4-Bromofluorobenzene	12.40	95	132702	51.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.48%	
Target Compounds						
20) Methylene Chloride	4.53	84	8527	3.047	ug/l #	83
43) Isopropyl Acetate	8.14	43	50393	11.538	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021020\
Data File : VN060050.D
Acq On : 10 Feb 2020 16:42
Operator : JC/MD
Sample : PB126661ZHE#17
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB126661ZHE#17

Quant Time: Feb 11 07:51:10 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
Quant Title : SW846 8260
QLast Update : Sat Jan 18 03:09:32 2020
Response via : Initial Calibration

