

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021021\
 Data File : VN065794.D
 Acq On : 10 Feb 2021 13:02
 Operator : JC/MD
 Sample : M1332-01
 Misc : 2.57g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TR-05-020521

Quant Time: Feb 10 13:23:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.618	168	160635	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.544	114	260559	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	236961	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.312	152	98273	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.981	65	88968	42.10	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	84.20%
35) Dibromofluoromethane	7.544	113	74260	48.09	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.18%
50) Toluene-d8	10.055	98	326399	53.78	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.56%
62) 4-Bromofluorobenzene	12.373	95	113294	51.84	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	103.68%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021021\
Data File : VN065794.D
Acq On : 10 Feb 2021 13:02
Operator : JC/MD
Sample : M1332-01
Misc : 2.57g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TR-05-020521

Quant Time: Feb 10 13:23:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

