



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021021\
 Data File : VN065790.D
 Acq On : 10 Feb 2021 9:56
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Title : SW846 8260
 Last Update : Mon Jan 25 13:47:28 2021



AutoFind: Scans 3305, 3306, 3307; Background Corrected with Scan 3293

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.8	4088	PASS
75	95	30	60	44.9	11661	PASS
95	95	100	100	100.0	25949	PASS
96	95	5	9	6.4	1656	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	74.0	19195	PASS
175	174	5	9	7.3	1410	PASS
176	174	95	101	99.2	19035	PASS
177	176	5	9	6.5	1246	PASS