

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021023\
 Data File : VN076607.D
 Acq On : 10 Feb 2023 14:20
 Operator : JC\MD
 Sample : 01457-02 10X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 14B-2

Quant Time: Feb 11 00:56:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N020923W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Feb 10 05:40:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.822	128	26232	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.104	114	182631	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.869	117	192017	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.586	65	95291	30.821	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	102.733%
60) 4-Bromofluorobenzene	12.851	95	101893	23.505	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	78.333%
63) Toluene-d8	10.569	98	200604	28.453	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	94.833%
Target Compounds						
					Qvalue	
25) Chloroform	7.975	83	67905	9.767	ug/l	84
41) Bromodichloromethane	9.892	83	7616	1.588	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021023\
Data File : VN076607.D
Acq On : 10 Feb 2023 14:20
Operator : JC\MD
Sample : 01457-02 10X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
14B-2

Quant Time: Feb 11 00:56:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N020923W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Feb 10 05:40:47 2023
Response via : Initial Calibration

