

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021023\
 Data File : VN076616.D
 Acq On : 10 Feb 2023 17:57
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Feb 11 00:58:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N020923W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Feb 10 05:40:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.822	128	27234	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.104	114	169381	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.869	117	179729	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.586	65	79901	24.892	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	82.967%#
60) 4-Bromofluorobenzene	12.851	95	93956	23.156	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	77.200%
63) Toluene-d8	10.569	98	186465	28.255	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	94.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021023\
Data File : VN076616.D
Acq On : 10 Feb 2023 17:57
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Feb 11 00:58:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N020923W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Feb 10 05:40:47 2023
Response via : Initial Calibration

