

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021023\
 Data File : VN076617.D
 Acq On : 10 Feb 2023 18:21
 Operator : JC\MD
 Sample : 01469-04RE 10X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 14B-4RE

Quant Time: Feb 11 00:59:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N020923W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Feb 10 05:40:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.822	128	26478	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.104	114	174547	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.869	117	188161	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.581	65	81892	26.241	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	87.467%#
60) 4-Bromofluorobenzene	12.851	95	94907	22.342	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	74.467%
63) Toluene-d8	10.569	98	190429	27.563	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	91.867%
Target Compounds						
25) Chloroform	7.975	83	49923	7.114	ug/l	Qvalue 95
41) Bromodichloromethane	9.892	83	6632	1.447	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021023\
Data File : VN076617.D
Acq On : 10 Feb 2023 18:21
Operator : JC\MD
Sample : 01469-04RE 10X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
14B-4RE

Quant Time: Feb 11 00:59:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N020923W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Feb 10 05:40:47 2023
Response via : Initial Calibration

