

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021025\
 Data File : VN085734.D
 Acq On : 10 Feb 2025 19:23
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Feb 11 03:29:17 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	259262	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	443928	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	390375	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	191076	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	185876	44.415	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.820%
35) Dibromofluoromethane	8.165	113	153775	49.931	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.860%
50) Toluene-d8	10.565	98	555929	50.805	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.620%
62) 4-Bromofluorobenzene	12.847	95	203792	54.445	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	108.880%

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.124	85	193680	55.174	ug/l	95
3) Chloromethane	2.359	50	179597	47.253	ug/l	98
4) Vinyl Chloride	2.512	62	196907	51.543	ug/l	98
5) Bromomethane	2.947	94	116700	50.572	ug/l	96
6) Chloroethane	3.118	64	118195	48.800	ug/l	92
7) Trichlorofluoromethane	3.494	101	275273	49.660	ug/l	100
8) Diethyl Ether	3.959	74	97506	50.919	ug/l	86
9) 1,1,2-Trichlorotrifluo...	4.371	101	154818	49.588	ug/l	98
10) Methyl Iodide	4.588	142	214003	59.855	ug/l	95
11) Tert butyl alcohol	5.518	59	137239	286.384	ug/l	99
12) 1,1-Dichloroethene	4.341	96	150353	54.035	ug/l	87
13) Acrolein	4.177	56	154428	236.067	ug/l	99
14) Allyl chloride	5.024	41	204071	45.198	ug/l	94
15) Acrylonitrile	5.712	53	384474	252.589	ug/l	99
16) Acetone	4.424	43	306483	226.652	ug/l	92
17) Carbon Disulfide	4.712	76	421833	49.238	ug/l	98
18) Methyl Acetate	5.018	43	192845	46.900	ug/l	94
19) Methyl tert-butyl Ether	5.794	73	494423	54.723	ug/l	96
20) Methylene Chloride	5.277	84	168393	50.304	ug/l	88
21) trans-1,2-Dichloroethene	5.788	96	155822	52.401	ug/l	94
22) Diisopropyl ether	6.671	45	479060	47.803	ug/l	94
23) Vinyl Acetate	6.600	43	1752867	249.932	ug/l	95
24) 1,1-Dichloroethane	6.565	63	297058	48.613	ug/l	97
25) 2-Butanone	7.477	43	487967	245.221	ug/l	92
26) 2,2-Dichloropropane	7.488	77	251850	50.977	ug/l	98
27) cis-1,2-Dichloroethene	7.482	96	185456	52.950	ug/l	94
28) Bromochloromethane	7.812	49	125852	44.268	ug/l	85
29) Tetrahydrofuran	7.835	42	323905	256.739	ug/l	91
30) Chloroform	7.965	83	308321	48.819	ug/l	99
31) Cyclohexane	8.253	56	238939	44.988	ug/l	92
32) 1,1,1-Trichloroethane	8.165	97	277255	50.047	ug/l	95
36) 1,1-Dichloropropene	8.371	75	215775	49.920	ug/l	98
37) Ethyl Acetate	7.553	43	197550	45.280	ug/l #	96
38) Carbon Tetrachloride	8.359	117	244270	49.363	ug/l	99
39) Methylcyclohexane	9.600	83	220053	54.177	ug/l	94
40) Benzene	8.606	78	668291	51.457	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021025\
 Data File : VN085734.D
 Acq On : 10 Feb 2025 19:23
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Feb 11 03:29:17 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	106931	47.110	ug/l	92
42) 1,2-Dichloroethane	8.671	62	223625	45.718	ug/l	98
43) Isopropyl Acetate	8.688	43	348408	49.838	ug/l	97
44) Trichloroethene	9.353	130	155378	51.396	ug/l	99
45) 1,2-Dichloropropane	9.618	63	162042	48.847	ug/l	100
46) Dibromomethane	9.706	93	116599	48.712	ug/l	94
47) Bromodichloromethane	9.882	83	244306	50.097	ug/l	100
48) Methyl methacrylate	9.676	41	156560	49.765	ug/l	92
49) 1,4-Dioxane	9.688	88	61182	1154.850	ug/l #	89
51) 4-Methyl-2-Pentanone	10.441	43	1023823	252.381	ug/l	94
52) Toluene	10.629	92	419396	55.732	ug/l	98
53) t-1,3-Dichloropropene	10.835	75	245908	53.359	ug/l	97
54) cis-1,3-Dichloropropene	10.312	75	265632	53.961	ug/l	92
55) 1,1,2-Trichloroethane	11.012	97	155717	52.288	ug/l	98
56) Ethyl methacrylate	10.870	69	251787	50.183	ug/l	88
57) 1,3-Dichloropropane	11.159	76	272119	52.550	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.159	63	594308	314.472	ug/l	97
59) 2-Hexanone	11.194	43	741839	259.894	ug/l	91
60) Dibromochloromethane	11.359	129	189230	52.632	ug/l	100
61) 1,2-Dibromoethane	11.465	107	156370	52.750	ug/l	100
64) Tetrachloroethene	11.100	164	141450	53.150	ug/l	95
65) Chlorobenzene	11.888	112	452748	52.942	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.959	131	163285	52.023	ug/l	99
67) Ethyl Benzene	11.959	91	783966	56.284	ug/l	97
68) m/p-Xylenes	12.064	106	609359	118.373	ug/l	96
69) o-Xylene	12.394	106	291664	59.286	ug/l	97
70) Styrene	12.412	104	487409	59.865	ug/l	98
71) Bromoform	12.576	173	124067	55.244	ug/l #	100
73) Isopropylbenzene	12.694	105	729301	56.552	ug/l	99
74) N-amyl acetate	12.494	43	286745	49.469	ug/l	94
75) 1,1,2,2-Tetrachloroethane	12.935	83	225447	49.490	ug/l	99
76) 1,2,3-Trichloropropane	12.988	75	290931	74.911	ug/l	46
77) Bromobenzene	12.976	156	178018	52.836	ug/l	95
78) n-propylbenzene	13.035	91	839211	54.970	ug/l	98
79) 2-Chlorotoluene	13.123	91	525231	53.158	ug/l	98
80) 1,3,5-Trimethylbenzene	13.170	105	601378	56.456	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	78030	54.360	ug/l #	81
82) 4-Chlorotoluene	13.217	91	522732	53.134	ug/l	98
83) tert-Butylbenzene	13.435	119	501874	56.104	ug/l	97
84) 1,2,4-Trimethylbenzene	13.476	105	606227	57.101	ug/l	98
85) sec-Butylbenzene	13.611	105	696881	56.209	ug/l	100
86) p-Isopropyltoluene	13.723	119	581123	51.819	ug/l	98
87) 1,3-Dichlorobenzene	13.729	146	319919	51.559	ug/l	99
88) 1,4-Dichlorobenzene	13.811	146	316227	49.169	ug/l	97
89) n-Butylbenzene	14.053	91	485026	54.531	ug/l	97
90) Hexachloroethane	14.329	117	110652	46.882	ug/l	95
91) 1,2-Dichlorobenzene	14.106	146	312433	50.480	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.717	75	43823	52.681	ug/l	91
93) 1,2,4-Trichlorobenzene	15.388	180	150156	52.199	ug/l	98
94) Hexachlorobutadiene	15.500	225	72862	47.709	ug/l	98
95) Naphthalene	15.635	128	493568	57.500	ug/l	100
96) 1,2,3-Trichlorobenzene	15.835	180	147693	50.742	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021025\
Data File : VN085734.D
Acq On : 10 Feb 2025 19:23
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :


Quant Time: Feb 11 03:29:17 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Feb 11 03:29:17 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.218 min Scan# 1065
Delta R.T. -0.000 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:168 Resp: 259262
Ion Ratio Lower Upper
168 100
99 63.9 53.6 80.4



#2
Dichlorodifluoromethane
Concen: 55.174 ug/l
RT: 2.124 min Scan# 29
Delta R.T. -0.000 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23

Tgt Ion: 85 Resp: 193680
Ion Ratio Lower Upper
85 100
87 30.7 16.7 50.1









#3
 Chloromethane
 Concen: 47.253 ug/l
 RT: 2.359 min Scan# 69
 Delta R.T. -0.000 min
 Lab File: VN085734.D
 Acq: 10 Feb 2025 19:23

Instrument :

MSVOA_N




ClientSampleId :


Tgt Ion: 50 Resp: 179597
 Ion Ratio Lower Upper
 50 100
 52 32.7 27.0 40.6



#4
 Vinyl Chloride
 Concen: 51.543 ug/l
 RT: 2.512 min Scan# 95
 Delta R.T. -0.000 min
 Lab File: VN085734.D
 Acq: 10 Feb 2025 19:23

Tgt Ion: 62 Resp: 196907
 Ion Ratio Lower Upper
 62 100
 64 31.9 24.8 37.2









#5
Bromomethane
Concen: 50.572 ug/l
RT: 2.947 min Scan# 11
Delta R.T. -0.006 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 94 Resp: 116700
Ion Ratio Lower Upper
94 100
96 93.3 71.8 107.6



#6
Chloroethane
Concen: 48.800 ug/l
RT: 3.118 min Scan# 198
Delta R.T. -0.000 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23


Tgt Ion: 64 Resp: 118195
Ion Ratio Lower Upper
64 100
66 35.1 24.6 36.8







#7
 Trichlorofluoromethane
 Concen: 49.660 ug/l
 RT: 3.494 min Scan# 21
 Delta R.T. -0.006 min
 Lab File: VN085734.D
 Acq: 10 Feb 2025 19:23

Instrument :
 MSVOA_N
 ClientSampleId :







Tgt Ion:101 Resp: 275273
 Ion Ratio Lower Upper
 101 100
 103 64.3 51.4 77.2



#8
 Diethyl Ether
 Concen: 50.919 ug/l
 RT: 3.959 min Scan# 341
 Delta R.T. 0.006 min
 Lab File: VN085734.D
 Acq: 10 Feb 2025 19:23

Tgt Ion: 74 Resp: 97506
 Ion Ratio Lower Upper
 74 100
 45 85.6 49.7 149.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 49.588 ug/l
RT: 4.371 min Scan# 411
Delta R.T. -0.000 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:101 Resp: 154818
Ion Ratio Lower Upper
101 100
85 44.8 37.8 56.8
151 74.2 58.8 88.2



#10
Methyl Iodide
Concen: 59.855 ug/l
RT: 4.588 min Scan# 448
Delta R.T. -0.000 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23

Tgt Ion:142 Resp: 214003
Ion Ratio Lower Upper
142 100
127 47.0 40.3 60.5
141 14.3 12.8 19.2









#11
Tert butyl alcohol
Concen: 286.384 ug/l
RT: 5.518 min Scan# 605
Delta R.T. 0.006 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 59 Resp: 137239
Ion Ratio Lower Upper
59 100
57 10.2 8.6 13.0



#12
1,1-Dichloroethene
Concen: 54.035 ug/l
RT: 4.341 min Scan# 406
Delta R.T. -0.000 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23

Tgt Ion: 96 Resp: 150353
Ion Ratio Lower Upper
96 100
61 154.4 141.0 211.4
98 64.3 49.0 73.6









#13
Acrolein
Concen: 236.067 ug/l
RT: 4.177 min Scan# 31
Delta R.T. 0.006 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23




Instrument :
MSVOA_N
ClientSampleId :

Tgt Ion: 56 Resp: 154428
Ion Ratio Lower Upper
56 100
55 69.9 56.3 84.5



#14
Allyl chloride
Concen: 45.198 ug/l
RT: 5.024 min Scan# 522
Delta R.T. 0.006 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23

Tgt Ion: 41 Resp: 204071
Ion Ratio Lower Upper
41 100
39 76.3 64.4 96.6
76 43.9 30.5 45.7









#15
Acrylonitrile
Concen: 252.589 ug/l
RT: 5.712 min Scan# 61
Delta R.T. -0.000 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 53 Resp: 384474
Ion Ratio Lower Upper
53 100
52 81.8 65.5 98.3
51 35.9 29.8 44.8



#16
Acetone
Concen: 226.652 ug/l
RT: 4.424 min Scan# 420
Delta R.T. 0.006 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23

Tgt Ion: 43 Resp: 306483
Ion Ratio Lower Upper
43 100
58 34.0 23.8 35.6









#17
Carbon Disulfide
Concen: 49.238 ug/l
RT: 4.712 min Scan# 41
Delta R.T. 0.006 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 76 Resp: 421833
Ion Ratio Lower Upper
76 100
78 9.4 6.9 10.3



#18
Methyl Acetate
Concen: 46.900 ug/l
RT: 5.018 min Scan# 521
Delta R.T. -0.000 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23

Tgt Ion: 43 Resp: 192845
Ion Ratio Lower Upper
43 100
74 26.4 18.6 28.0









#19
Methyl tert-butyl Ether
Concen: 54.723 ug/l
RT: 5.794 min Scan# 61
Delta R.T. 0.006 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 73 Resp: 494423
Ion Ratio Lower Upper
73 100
57 20.6 17.8 26.8



#20
Methylene Chloride
Concen: 50.304 ug/l
RT: 5.277 min Scan# 565
Delta R.T. 0.006 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23

Tgt Ion: 84 Resp: 168393
Ion Ratio Lower Upper
84 100
49 107.6 104.1 156.1
51 36.4 31.0 46.4
86 62.9 51.9 77.9









#21
trans-1,2-Dichloroethene
Concen: 52.401 ug/l
RT: 5.788 min Scan# 61
Delta R.T. 0.006 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23




Instrument :
MSVOA_N
ClientSampleId :

Tgt Ion: 96 Resp: 155822
Ion Ratio Lower Upper
96 100
61 134.4 115.7 173.5
98 63.2 49.8 74.8



#22
Diisopropyl ether
Concen: 47.803 ug/l
RT: 6.671 min Scan# 802
Delta R.T. 0.006 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23

Tgt Ion: 45 Resp: 479060
Ion Ratio Lower Upper
45 100
43 52.4 45.0 67.6
87 31.4 21.8 32.8
59 12.0 10.2 15.2









#23
 Vinyl Acetate
 Concen: 249.932 ug/l
 RT: 6.600 min Scan# 790
 Delta R.T. 0.006 min
 Lab File: VN085734.D
 Acq: 10 Feb 2025 19:23




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion: 43 Resp: 1752867
 Ion Ratio Lower Upper
 43 100
 86 12.4 8.4 12.6



#24
 1,1-Dichloroethane
 Concen: 48.613 ug/l
 RT: 6.565 min Scan# 784
 Delta R.T. -0.000 min
 Lab File: VN085734.D
 Acq: 10 Feb 2025 19:23

Tgt Ion: 63 Resp: 297058
 Ion Ratio Lower Upper
 63 100
 98 7.3 3.3 9.8
 100 4.9 2.0 6.0









#25
2-Butanone
Concen: 245.221 ug/l
RT: 7.477 min Scan# 91
Delta R.T. -0.000 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 43 Resp: 487967
Ion Ratio Lower Upper
43 100
72 29.5 20.2 30.4



#26
2,2-Dichloropropane
Concen: 50.977 ug/l
RT: 7.488 min Scan# 941
Delta R.T. 0.006 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23

Tgt Ion: 77 Resp: 251850
Ion Ratio Lower Upper
77 100
97 22.5 10.7 32.1









#27
 cis-1,2-Dichloroethene
 Concen: 52.950 ug/l
 RT: 7.482 min Scan# 940
 Delta R.T. -0.000 min
 Lab File: VN085734.D
 Acq: 10 Feb 2025 19:23




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion: 96 Resp: 185456
 Ion Ratio Lower Upper
 96 100
 61 145.8 0.0 311.8
 98 65.2 0.0 126.0



#28
 Bromochloromethane
 Concen: 44.268 ug/l
 RT: 7.812 min Scan# 996
 Delta R.T. 0.006 min
 Lab File: VN085734.D
 Acq: 10 Feb 2025 19:23

Tgt Ion: 49 Resp: 125852
 Ion Ratio Lower Upper
 49 100
 129 2.0 0.0 4.4
 130 81.4 55.0 82.4





#29

Tetrahydrofuran

Concen: 256.739 ug/l

RT: 7.835 min Scan# 1000

Delta R.T. 0.006 min

Lab File: VN085734.D

Acq: 10 Feb 2025 19:23

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion: 42 Resp: 323905

Ion Ratio Lower Upper

42 100

72 49.3 34.2 51.2

71 45.9 32.5 48.7



#30

Chloroform

Concen: 48.819 ug/l

RT: 7.965 min Scan# 1022

Delta R.T. 0.006 min

Lab File: VN085734.D




Acq: 10 Feb 2025 19:23


Tgt Ion: 83 Resp: 308321

Ion Ratio Lower Upper

83 100

85 63.7 51.8 77.6









#31
Cyclohexane
Concen: 44.988 ug/l
RT: 8.253 min Scan# 1071
Delta R.T. -0.000 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 56 Resp: 238939
Ion Ratio Lower Upper
56 100
69 31.8 22.2 33.4
84 90.4 66.4 99.6



#32
1,1,1-Trichloroethane
Concen: 50.047 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. -0.000 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23

Tgt Ion: 97 Resp: 277255
Ion Ratio Lower Upper
97 100
99 64.6 49.8 74.6
61 46.4 41.4 62.2





#33

1,2-Dichloroethane-d4

Concen: 44.415 ug/l

RT: 8.576 min Scan# 1125

Delta R.T. 0.006 min

Lab File: VN085734.D

Acq: 10 Feb 2025 19:23

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 65 Resp: 185876

Ion Ratio Lower Upper

65 100

67 52.0 0.0 101.6



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1215

Delta R.T. 0.006 min

Lab File: VN085734.D

Acq: 10 Feb 2025 19:23




Tgt Ion:114 Resp: 443928

Ion Ratio Lower Upper

114 100

63 20.5 0.0 47.6

88 15.8 0.0 32.6





#35

Dibromofluoromethane

Concen: 49.931 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. 0.006 min

Lab File: VN085734.D

Acq: 10 Feb 2025 19:23

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion:113 Resp: 153775

Ion Ratio Lower Upper

113 100

111 103.4 82.7 124.1

192 18.3 14.3 21.5



#36

1,1-Dichloropropene

Concen: 49.920 ug/l

RT: 8.371 min Scan# 1091

Delta R.T. 0.006 min

Lab File: VN085734.D

Acq: 10 Feb 2025 19:23




Tgt Ion: 75 Resp: 215775


Ion Ratio Lower Upper

75 100

110 34.3 16.5 49.5

77 31.6 24.4 36.6





#37

Ethyl Acetate

Concen: 45.280 ug/l

RT: 7.553 min Scan# 91

Delta R.T. -0.000 min

Lab File: VN085734.D

Acq: 10 Feb 2025 19:23

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion: 43 Resp: 197550

Ion Ratio Lower Upper

43 100

61 14.3 10.9 16.3

70 11.7 7.5 11.3#



#38

Carbon Tetrachloride

Concen: 49.363 ug/l

RT: 8.359 min Scan# 1089

Delta R.T. -0.000 min

Lab File: VN085734.D

Acq: 10 Feb 2025 19:23




Tgt Ion: 117 Resp: 244270


Ion Ratio Lower Upper

117 100

119 95.8 75.4 113.2

121 31.4 24.6 37.0









#39
Methylcyclohexane
Concen: 54.177 ug/l
RT: 9.600 min Scan# 1130
Delta R.T. -0.000 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 83 Resp: 220053
Ion Ratio Lower Upper
83 100
55 71.1 62.6 94.0
98 48.6 37.7 56.5



#40
Benzene
Concen: 51.457 ug/l
RT: 8.606 min Scan# 1131
Delta R.T. 0.006 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23

Tgt Ion: 78 Resp: 668291
Ion Ratio Lower Upper
78 100
77 24.4 19.0 28.6









#41
Methacrylonitrile
Concen: 47.110 ug/l
RT: 7.777 min Scan# 91
Delta R.T. 0.006 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23

Instrument :
MSVOA_N
ClientSampleId :

Tgt Ion: 41 Resp: 106931




Ion	Ratio	Lower	Upper
41	100		
39	53.9	46.0	69.0
67	82.3	57.4	86.2
52	33.7	25.5	38.3




#42
1,2-Dichloroethane
Concen: 45.718 ug/l
RT: 8.671 min Scan# 1142
Delta R.T. 0.006 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23

Tgt Ion: 62 Resp: 223625

Ion	Ratio	Lower	Upper
62	100		
98	9.2	0.0	17.0









#43
Isopropyl Acetate
Concen: 49.838 ug/l
RT: 8.688 min Scan# 1145
Delta R.T. 0.006 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 43 Resp: 348408
Ion Ratio Lower Upper
43 100
61 27.3 20.7 31.1
87 13.9 9.8 14.8



#44
Trichloroethene
Concen: 51.396 ug/l
RT: 9.353 min Scan# 1258
Delta R.T. 0.006 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23

Tgt Ion:130 Resp: 155378
Ion Ratio Lower Upper
130 100
95 96.7 0.0 195.8





#45

1,2-Dichloropropane

Concen: 48.847 ug/l

RT: 9.618 min Scan# 11

Delta R.T. -0.000 min

Lab File: VN085734.D

Acq: 10 Feb 2025 19:23

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 63 Resp: 162042

Ion Ratio Lower Upper

63 100

65 32.1 25.6 38.4



#46

Dibromomethane

Concen: 48.712 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. -0.000 min

Lab File: VN085734.D

Acq: 10 Feb 2025 19:23




Tgt Ion: 93 Resp: 116599


Ion Ratio Lower Upper

93 100

95 84.0 64.7 97.1

174 93.0 69.0 103.4









#47
Bromodichloromethane
Concen: 50.097 ug/l
RT: 9.882 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 83 Resp: 244306
Ion Ratio Lower Upper
83 100
85 64.1 51.2 76.8
127 8.6 6.5 9.7



#48
Methyl methacrylate
Concen: 49.765 ug/l
RT: 9.676 min Scan# 1313
Delta R.T. -0.001 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23

Tgt Ion: 41 Resp: 156560
Ion Ratio Lower Upper
41 100
69 91.8 64.7 97.1
39 59.7 49.0 73.6





#49

1,4-Dioxane

Concen: 1154.850 ug/l

RT: 9.688 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN085734.D

Acq: 10 Feb 2025 19:23

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion: 88 Resp: 61182

Ion Ratio Lower Upper

88 100

43 24.2 26.6 39.8#

58 66.9 59.5 89.3



#50

Toluene-d8

Concen: 50.805 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. -0.000 min

Lab File: VN085734.D




Acq: 10 Feb 2025 19:23


Tgt Ion: 98 Resp: 555929

Ion Ratio Lower Upper

98 100

100 64.7 52.2 78.4









#51
4-Methyl-2-Pentanone
Concen: 252.381 ug/l
RT: 10.441 min Scan# 1443
Delta R.T. 0.000 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 43 Resp: 1023823
Ion Ratio Lower Upper
43 100
58 42.0 30.5 45.7



#52
Toluene
Concen: 55.732 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. 0.005 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23

Tgt Ion: 92 Resp: 419396
Ion Ratio Lower Upper
92 100
91 171.4 139.2 208.8





#53

t-1,3-Dichloropropene

Concen: 53.359 ug/l

RT: 10.835 min Scan# 11

Delta R.T. -0.000 min

Lab File: VN085734.D

Acq: 10 Feb 2025 19:23

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 75 Resp: 245908

Ion Ratio Lower Upper

75 100

77 31.2 23.5 35.3



#54

cis-1,3-Dichloropropene

Concen: 53.961 ug/l

RT: 10.312 min Scan# 1421

Delta R.T. 0.006 min

Lab File: VN085734.D

Acq: 10 Feb 2025 19:23




Tgt Ion: 75 Resp: 265632


Ion Ratio Lower Upper

75 100

77 31.3 25.0 37.4

39 44.5 43.1 64.7









#55
 1,1,2-Trichloroethane
 Concen: 52.288 ug/l
 RT: 11.012 min Scan# 1540
 Delta R.T. -0.000 min
 Lab File: VN085734.D
 Acq: 10 Feb 2025 19:23




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion: 97	Resp: 155717
Ion Ratio	Lower Upper
97 100	
83 85.4	69.0 103.6
85 58.0	44.6 66.8
99 62.5	48.6 72.8



#56
 Ethyl methacrylate
 Concen: 50.183 ug/l
 RT: 10.870 min Scan# 1516
 Delta R.T. -0.001 min
 Lab File: VN085734.D
 Acq: 10 Feb 2025 19:23

Tgt Ion: 69	Resp: 251787
Ion Ratio	Lower Upper
69 100	
41 58.6	54.6 82.0
39 32.4	32.4 48.6





#57

1,3-Dichloropropane

Concen: 52.550 ug/l

RT: 11.159 min Scan# 11

Delta R.T. -0.000 min

Lab File: VN085734.D

Acq: 10 Feb 2025 19:23

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 76 Resp: 272119

Ion Ratio Lower Upper

76 100

78 32.0 25.6 38.4



#58

2-Chloroethyl Vinyl ether

Concen: 314.472 ug/l

RT: 10.159 min Scan# 1395

Delta R.T. -0.000 min

Lab File: VN085734.D




Acq: 10 Feb 2025 19:23


Tgt Ion: 63 Resp: 594308

Ion Ratio Lower Upper

63 100

106 28.6 21.6 32.4









#59
2-Hexanone
Concen: 259.894 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. 0.006 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 43 Resp: 741839
Ion Ratio Lower Upper
43 100
58 58.5 26.2 78.6



#60
Dibromochloromethane
Concen: 52.632 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. 0.006 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23

Tgt Ion: 129 Resp: 189230
Ion Ratio Lower Upper
129 100
127 77.3 38.6 115.8









#61
1,2-Dibromoethane
Concen: 52.750 ug/l
RT: 11.465 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:107 Resp: 156370
Ion Ratio Lower Upper
107 100
109 94.6 75.9 113.9



#62
4-Bromofluorobenzene
Concen: 54.445 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23

Tgt Ion: 95 Resp: 203792
Ion Ratio Lower Upper
95 100
174 75.2 0.0 145.0
176 71.8 0.0 142.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 1170
Delta R.T. -0.000 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:117 Resp: 390375
Ion Ratio Lower Upper
117 100
82 56.3 48.6 72.8
119 32.7 26.6 39.8



#64
Tetrachloroethene
Concen: 53.150 ug/l
RT: 11.100 min Scan# 1555
Delta R.T. -0.000 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23

Tgt Ion:164 Resp: 141450
Ion Ratio Lower Upper
164 100
166 122.5 103.4 155.2
129 94.1 79.2 118.8
131 92.1 77.1 115.7









#65
Chlorobenzene
Concen: 52.942 ug/l
RT: 11.888 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:112 Resp: 452748
Ion Ratio Lower Upper
112 100
114 32.0 25.3 37.9



#66
1,1,1,2-Tetrachloroethane
Concen: 52.023 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23

Tgt Ion:131 Resp: 163285
Ion Ratio Lower Upper
131 100
133 94.9 47.4 142.3
119 64.8 33.1 99.5









#67
Ethyl Benzene
Concen: 56.284 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 91 Resp: 783966
Ion Ratio Lower Upper
91 100
106 31.6 23.8 35.8



#68
m/p-Xylenes
Concen: 118.373 ug/l
RT: 12.064 min Scan# 1719
Delta R.T. -0.000 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23

Tgt Ion:106 Resp: 609359
Ion Ratio Lower Upper
106 100
91 203.1 167.7 251.5









#69
o-Xylene
Concen: 59.286 ug/l
RT: 12.394 min Scan# 1775
Delta R.T. -0.000 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:106 Resp: 291664
Ion Ratio Lower Upper
106 100
91 215.6 110.4 331.2



#70
Styrene
Concen: 59.865 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. 0.006 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23

Tgt Ion:104 Resp: 487409
Ion Ratio Lower Upper
104 100
78 50.7 42.5 63.7
103 55.1 43.8 65.8









#71
 Bromoform
 Concen: 55.244 ug/l
 RT: 12.576 min Scan# 1806
 Delta R.T. -0.000 min
 Lab File: VN085734.D
 Acq: 10 Feb 2025 19:23




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion:	173	Resp:	124067
Ion Ratio	Lower	Upper	
173	100		
175	48.8	24.4	73.2
254	0.0	0.0	0.0



#72
 1,4-Dichlorobenzene-d4
 Concen: 50.000 ug/l
 RT: 13.788 min Scan# 2012
 Delta R.T. -0.000 min
 Lab File: VN085734.D
 Acq: 10 Feb 2025 19:23

Tgt Ion:	152	Resp:	191076
Ion Ratio	Lower	Upper	
152	100		
115	60.3	31.1	93.3
150	173.4	0.0	343.6









#73
Isopropylbenzene
Concen: 56.552 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. 0.006 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:105 Resp: 729301
Ion Ratio Lower Upper
105 100
120 26.2 12.8 38.3



#74
N-ethyl acetate
Concen: 49.469 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. 0.006 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23

Tgt Ion: 43 Resp: 286745
Ion Ratio Lower Upper
43 100
70 47.9 34.0 51.0
55 27.6 21.4 32.2
61 26.4 19.0 28.4









#75
 1,1,2,2-Tetrachloroethane
 Concen: 49.490 ug/l
 RT: 12.935 min Scan# 1867
 Delta R.T. -0.000 min
 Lab File: VN085734.D
 Acq: 10 Feb 2025 19:23




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion: 83 Resp: 225447
 Ion Ratio Lower Upper
 83 100
 131 10.6 4.8 14.4
 85 64.2 32.2 96.6



#76
 1,2,3-Trichloropropane
 Concen: 74.911 ug/l
 RT: 12.988 min Scan# 1876
 Delta R.T. -0.000 min
 Lab File: VN085734.D
 Acq: 10 Feb 2025 19:23

Tgt Ion: 75 Resp: 290931
 Ion Ratio Lower Upper
 75 100
 77 132.5 109.7 329.2









#77
Bromobenzene
Concen: 52.836 ug/l
RT: 12.976 min Scan# 1874
Delta R.T. -0.000 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:156 Resp: 178018
Ion Ratio Lower Upper
156 100
77 216.6 114.1 342.4
158 96.6 48.9 146.8



#78
n-propylbenzene
Concen: 54.970 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. 0.006 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23

Tgt Ion: 91 Resp: 839211
Ion Ratio Lower Upper
91 100
120 22.5 10.9 32.6









#79
2-Chlorotoluene
Concen: 53.158 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. -0.000 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 91 Resp: 525231
Ion Ratio Lower Upper
91 100
126 32.6 15.7 47.1



#80
1,3,5-Trimethylbenzene
Concen: 56.456 ug/l
RT: 13.170 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23

Tgt Ion: 105 Resp: 601378
Ion Ratio Lower Upper
105 100
120 47.6 23.9 71.7









#81
trans-1,4-Dichloro-2-butene
Concen: 54.360 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. -0.000 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 75 Resp: 78030
Ion Ratio Lower Upper
75 100
53 91.7 95.6 143.4#
89 47.3 37.0 55.6



#82
4-Chlorotoluene
Concen: 53.134 ug/l
RT: 13.217 min Scan# 1915
Delta R.T. -0.000 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23

Tgt Ion: 91 Resp: 522732
Ion Ratio Lower Upper
91 100
126 32.7 15.9 47.7





#83

tert-Butylbenzene

Concen: 56.104 ug/l

RT: 13.435 min Scan# 1952

Delta R.T. -0.000 min

Lab File: VN085734.D

Acq: 10 Feb 2025 19:23

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion:119 Resp: 501874

Ion Ratio Lower Upper

119 100

91 67.5 35.1 105.3

134 24.6 12.0 36.1



#84

1,2,4-Trimethylbenzene

Concen: 57.101 ug/l

RT: 13.476 min Scan# 1959

Delta R.T. -0.000 min

Lab File: VN085734.D




Acq: 10 Feb 2025 19:23


Tgt Ion:105 Resp: 606227

Ion Ratio Lower Upper

105 100

120 44.9 21.6 65.0









#85
 sec-Butylbenzene
 Concen: 56.209 ug/l
 RT: 13.611 min Scan# 1982
 Delta R.T. -0.001 min
 Lab File: VN085734.D
 Acq: 10 Feb 2025 19:23




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion:105 Resp: 696881
 Ion Ratio Lower Upper
 105 100
 134 19.2 9.7 28.9



#86
 p-Isopropyltoluene
 Concen: 51.819 ug/l
 RT: 13.723 min Scan# 2001
 Delta R.T. 0.000 min
 Lab File: VN085734.D
 Acq: 10 Feb 2025 19:23

Tgt Ion:119 Resp: 581123
 Ion Ratio Lower Upper
 119 100
 134 25.5 12.7 38.0
 91 23.7 12.7 38.1









#87
 1,3-Dichlorobenzene
 Concen: 51.559 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. -0.000 min
 Lab File: VN085734.D
 Acq: 10 Feb 2025 19:23




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion:146 Resp: 319919
 Ion Ratio Lower Upper
 146 100
 111 42.0 21.4 64.3
 148 63.6 32.3 96.9



#88
 1,4-Dichlorobenzene
 Concen: 49.169 ug/l
 RT: 13.811 min Scan# 2016
 Delta R.T. -0.000 min
 Lab File: VN085734.D
 Acq: 10 Feb 2025 19:23

Tgt Ion:146 Resp: 316227
 Ion Ratio Lower Upper
 146 100
 111 40.0 21.3 63.7
 148 63.3 32.4 97.0





#89

n-Butylbenzene

Concen: 54.531 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. -0.000 min

Lab File: VN085734.D

Acq: 10 Feb 2025 19:23

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion: 91 Resp: 485026

Ion Ratio Lower Upper

91 100

92 53.0 25.8 77.3

134 25.3 11.7 35.1



#90

Hexachloroethane

Concen: 46.882 ug/l

RT: 14.329 min Scan# 2104

Delta R.T. -0.000 min

Lab File: VN085734.D




Acq: 10 Feb 2025 19:23


Tgt Ion: 117 Resp: 110652

Ion Ratio Lower Upper

117 100


201 71.3 33.7 101.0







#91
1,2-Dichlorobenzene
Concen: 50.480 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. 0.006 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion:146 Resp: 312433
Ion Ratio Lower Upper
146 100
111 42.9 21.7 65.1
148 62.9 31.4 94.2



#92
1,2-Dibromo-3-Chloropropane
Concen: 52.681 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. -0.000 min
Lab File: VN085734.D
Acq: 10 Feb 2025 19:23

Tgt Ion: 75 Resp: 43823
Ion Ratio Lower Upper
75 100
155 78.4 36.4 109.2
157 101.0 45.4 136.1





#93

1,2,4-Trichlorobenzene

Concen: 52.199 ug/l

RT: 15.388 min Scan# 2284

Delta R.T. -0.000 min

Lab File: VN085734.D

Acq: 10 Feb 2025 19:23

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion:180 Resp: 150156

Ion Ratio Lower Upper

180 100

182 95.8 46.8 140.3

145 32.9 16.0 48.0



#94

Hexachlorobutadiene

Concen: 47.709 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. -0.000 min

Lab File: VN085734.D

Acq: 10 Feb 2025 19:23




Tgt Ion:225 Resp: 72862


Ion Ratio Lower Upper

225 100

223 63.9 30.7 92.1

227 62.2 30.9 92.5





#95

Naphthalene

Concen: 57.500 ug/l

RT: 15.635 min Scan# 2326

Delta R.T. -0.000 min

Lab File: VN085734.D

Acq: 10 Feb 2025 19:23

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion:128 Resp: 493568

Ion Ratio Lower Upper

128 100

127 13.2 10.6 16.0

129 11.0 8.8 13.2



#96

1,2,3-Trichlorobenzene

Concen: 50.742 ug/l

RT: 15.835 min Scan# 2360

Delta R.T. -0.000 min

Lab File: VN085734.D

Acq: 10 Feb 2025 19:23

Tgt Ion:180 Resp: 147693

Ion Ratio Lower Upper

180 100

182 93.9 47.4 142.2

145 34.0 16.9 50.7

