

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021025\
 Data File : VN085728.D
 Acq On : 10 Feb 2025 16:59
 Operator : JC\MD
 Sample : Q1328-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 Storage-Blank-WATER-Ref-5

Quant Time: Feb 11 03:26:09 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	229831	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	425415	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	373084	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	146687	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	183516	49.466	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.940%
35) Dibromofluoromethane	8.165	113	150571	51.018	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.040%
50) Toluene-d8	10.565	98	517689	49.369	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.740%
62) 4-Bromofluorobenzene	12.847	95	164897	45.971	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	91.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021025\
Data File : VN085728.D
Acq On : 10 Feb 2025 16:59
Operator : JC\MD
Sample : Q1328-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
Storage-Blank-WATER-Ref-5

Quant Time: Feb 11 03:26:09 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

