

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021119\
 Data File : VN053697.D
 Acq On : 11 Feb 2019 11:52
 Operator : JC/SP
 Sample : PB116972TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB116972TB

Quant Time: Feb 11 22:28:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	733849	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1227211	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1114112	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	436447	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	396691	53.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.76%	
35) Dibromofluoromethane	7.59	113	396170	51.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.18%	
50) Toluene-d8	10.09	98	1495192	51.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.70%	
62) 4-Bromofluorobenzene	12.40	95	481369	49.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.14%	
Target Compounds						
16) Acetone	3.82	43	19089	10.109	ug/l	96
20) Methylene Chloride	4.55	84	17605	2.222	ug/l	98
25) 2-Butanone	6.85	43	3328	1.236	ug/l	90
43) Isopropyl Acetate	8.17	43	26834	2.249	ug/l #	85
95) Naphthalene	15.13	128	5629	2.020	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021119\
Data File : VN053697.D
Acq On : 11 Feb 2019 11:52
Operator : JC/SP
Sample : PB116972TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB116972TB

Quant Time: Feb 11 22:28:17 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

