

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021119\
 Data File : VN053701.D
 Acq On : 11 Feb 2019 13:39
 Operator : JC/SP
 Sample : K1356-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SU-03-020819-B

Quant Time: Feb 11 22:47:08 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	756627	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1253662	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1173269	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	514821	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	402826	53.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.14%	
35) Dibromofluoromethane	7.59	113	394521	49.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.60%	
50) Toluene-d8	10.09	98	1525971	51.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.60%	
62) 4-Bromofluorobenzene	12.40	95	526390	53.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.12%	
Target Compounds						
16) Acetone	3.82	43	16517	8.483	ug/l	95
20) Methylene Chloride	4.55	84	25761	3.154	ug/l	93
43) Isopropyl Acetate	8.17	43	32749	2.687	ug/l #	82
95) Naphthalene	15.13	128	480376	22.240	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021119\
Data File : VN053701.D
Acq On : 11 Feb 2019 13:39
Operator : JC/SP
Sample : K1356-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SU-03-020819-B

Quant Time: Feb 11 22:47:08 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

