

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021119\
 Data File : VN053705.D
 Acq On : 11 Feb 2019 15:26
 Operator : JC/SP
 Sample : K1442-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 COMP-11

Quant Time: Feb 11 22:59:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	757694	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1242117	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1138905	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	448891	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	402091	52.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.80%	
35) Dibromofluoromethane	7.59	113	393774	50.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.34%	
50) Toluene-d8	10.09	98	1522633	51.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.34%	
62) 4-Bromofluorobenzene	12.40	95	488126	49.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.32%	
Target Compounds						
16) Acetone	3.82	43	17089	8.765	ug/l	97
20) Methylene Chloride	4.55	84	9225964	1127.893	ug/l	99
43) Isopropyl Acetate	8.17	43	33923	2.809	ug/l #	74
95) Naphthalene	15.13	128	12331	2.340	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021119\
Data File : VN053705.D
Acq On : 11 Feb 2019 15:26
Operator : JC/SP
Sample : K1442-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
COMP-11

Quant Time: Feb 11 22:59:15 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

