

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021119\
 Data File : VN053706.D
 Acq On : 11 Feb 2019 15:53
 Operator : JC/SP
 Sample : K1450-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SHORE-RD-BOX-CULVERT-TP

Quant Time: Feb 11 23:02:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	747340	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1198779	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1091570	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	432501	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	387488	51.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.36%	
35) Dibromofluoromethane	7.59	113	379477	50.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.20%	
50) Toluene-d8	10.09	98	1458270	51.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.54%	
62) 4-Bromofluorobenzene	12.40	95	470360	49.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.18%	
Target Compounds						
16) Acetone	3.82	43	23101	12.013	ug/l	96
20) Methylene Chloride	4.55	84	45598	5.652	ug/l	98
40) Benzene	8.05	78	45305	1.360	ug/l	98
43) Isopropyl Acetate	8.17	43	34956	2.999	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021119\
Data File : VN053706.D
Acq On : 11 Feb 2019 15:53
Operator : JC/SP
Sample : K1450-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SHORE-RD-BOX-CULVERT-TP

Quant Time: Feb 11 23:02:27 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

