

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021119\
 Data File : VN053707.D
 Acq On : 11 Feb 2019 16:20
 Operator : JC/SP
 Sample : K1450-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 HOSPITAL-BOX-CULVERT-TP

Quant Time: Feb 11 23:03:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	729708	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1196807	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1085168	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	414750	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	387465	52.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.86%	
35) Dibromofluoromethane	7.59	113	375281	49.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.26%	
50) Toluene-d8	10.09	98	1453991	51.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.42%	
62) 4-Bromofluorobenzene	12.40	95	459611	48.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.06%	
Target Compounds						
16) Acetone	3.82	43	27201	14.486	ug/l	99
20) Methylene Chloride	4.55	84	36469	4.629	ug/l	98
25) 2-Butanone	6.85	43	5885	2.198	ug/l	98
43) Isopropyl Acetate	8.17	43	32864	2.824	ug/l #	78
95) Naphthalene	15.13	128	83399	6.155	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021119\
Data File : VN053707.D
Acq On : 11 Feb 2019 16:20
Operator : JC/SP
Sample : K1450-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
HOSPITAL-BOX-CULVERT-TP

Quant Time: Feb 11 23:03:48 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

