

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021119\
 Data File : VN053708.D
 Acq On : 11 Feb 2019 16:47
 Operator : JC/SP
 Sample : K1453-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 M-8-S

Quant Time: Feb 11 23:06:59 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	729509	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1189777	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1089550	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	435760	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	383402	52.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.78%	
35) Dibromofluoromethane	7.59	113	382399	50.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.74%	
50) Toluene-d8	10.09	98	1447665	51.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.56%	
62) 4-Bromofluorobenzene	12.40	95	472478	50.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.38%	
Target Compounds						
16) Acetone	3.82	43	12058	6.423	ug/l	90
20) Methylene Chloride	4.55	84	4827564	612.981	ug/l	99
43) Isopropyl Acetate	8.17	43	34537	2.986	ug/l #	80
95) Naphthalene	15.14	128	17526	2.620	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021119\
Data File : VN053708.D
Acq On : 11 Feb 2019 16:47
Operator : JC/SP
Sample : K1453-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
M-8-S

Quant Time: Feb 11 23:06:59 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

