

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021119\
 Data File : VN053710.D
 Acq On : 11 Feb 2019 17:40
 Operator : JC/SP
 Sample : K1453-12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 M-9-S

Quant Time: Feb 11 23:10:26 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	720469	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1179208	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1066563	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	424013	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	382721	52.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.90%	
35) Dibromofluoromethane	7.59	113	378894	50.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.70%	
50) Toluene-d8	10.09	98	1423661	50.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.78%	
62) 4-Bromofluorobenzene	12.40	95	461033	49.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.82%	
Target Compounds						
16) Acetone	3.82	43	14944	8.061	ug/l	100
20) Methylene Chloride	4.55	84	324328	41.698	ug/l	98
43) Isopropyl Acetate	8.17	43	38696	3.375	ug/l #	79
95) Naphthalene	15.13	128	23133	2.935	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021119\
Data File : VN053710.D
Acq On : 11 Feb 2019 17:40
Operator : JC/SP
Sample : K1453-12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
M-9-S

Quant Time: Feb 11 23:10:26 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

