

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN021119\  
 Data File : VN053694.D  
 Acq On : 11 Feb 2019 9:57  
 Operator : JC/SP  
 Sample : VSTDCCC050  
 Misc : 5.00mL/MSVOA N/WATER  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 LabSampleId :  
 VSTDCCC050

Quant Time: Feb 11 22:22:08 2019  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_N\METHODS\82N013019W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Feb 01 09:11:52 2019  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|       | Compound                    | AvqRF | CCRF  | %Dev   | Area% | Dev(min) |
|-------|-----------------------------|-------|-------|--------|-------|----------|
| 1 I   | Pentafluorobenzene          | 1.000 | 1.000 | 0.0    | 103   | 0.00     |
| 2 T   | Dichlorodifluoromethane     | 0.515 | 0.534 | -3.7   | 98    | 0.00     |
| 3 P   | Chloromethane               | 0.594 | 0.558 | 6.1    | 97    | 0.00     |
| 4 C   | Vinyl Chloride              | 0.576 | 0.565 | 1.9#   | 100   | 0.00     |
| 5 T   | Bromomethane                | 0.377 | 0.393 | -4.2   | 110   | 0.00     |
| 6 T   | Chloroethane                | 0.329 | 0.359 | -9.1   | 114   | 0.00     |
| 7 T   | Trichlorofluoromethane      | 0.781 | 0.781 | 0.0    | 104   | 0.00     |
| 8 T   | Diethyl Ether               | 0.285 | 0.275 | 3.5    | 101   | 0.00     |
| 9 T   | 1,1,2-Trichlorotrifluoroeth | 0.483 | 0.492 | -1.9   | 107   | 0.00     |
| 10 T  | Methyl Iodide               | 0.759 | 0.722 | 4.9    | 99    | 0.00     |
| 11 T  | Tert butyl alcohol          | 0.030 | 0.028 | 6.7    | 100   | 0.00     |
| 12 CM | 1,1-Dichloroethene          | 0.463 | 0.457 | 1.3#   | 102   | 0.00     |
| 13 T  | Acrolein                    | 0.016 | 0.010 | 37.5#  | 82    | 0.00     |
| 14 T  | Allyl chloride              | 0.737 | 0.734 | 0.4    | 102   | 0.00     |
| 15 T  | Acrylonitrile               | 0.160 | 0.161 | -0.6   | 103   | 0.00     |
| 16 T  | Acetone                     | 0.129 | 0.166 | -28.7# | 143   | 0.00     |
| 17 T  | Carbon Disulfide            | 1.442 | 1.317 | 8.7    | 96    | 0.00     |
| 18 T  | Methyl Acetate              | 0.381 | 0.389 | -2.1   | 111   | 0.00     |
| 19 T  | Methyl tert-butyl Ether     | 1.174 | 1.192 | -1.5   | 103   | 0.00     |
| 20 T  | Methylene Chloride          | 0.540 | 0.521 | 3.5    | 106   | 0.00     |
| 21 T  | trans-1,2-Dichloroethene    | 0.500 | 0.486 | 2.8    | 102   | 0.00     |
| 22 T  | Diisopropyl ether           | 1.366 | 1.421 | -4.0   | 105   | 0.00     |
| 23 T  | Vinyl Acetate               | 0.945 | 0.941 | 0.4    | 98    | 0.00     |
| 24 P  | 1,1-Dichloroethane          | 0.911 | 0.906 | 0.5    | 105   | 0.00     |
| 25 T  | 2-Butanone                  | 0.183 | 0.194 | -6.0   | 109   | 0.00     |
| 26 T  | 2,2-Dichloropropane         | 0.696 | 0.703 | -1.0   | 106   | 0.00     |
| 27 T  | cis-1,2-Dichloroethene      | 0.555 | 0.557 | -0.4   | 103   | 0.00     |
| 28 T  | Bromochloromethane          | 0.395 | 0.404 | -2.3   | 107   | 0.00     |
| 29 T  | Tetrahydrofuran             | 0.118 | 0.115 | 2.5    | 99    | 0.00     |
| 30 C  | Chloroform                  | 0.903 | 0.891 | 1.3#   | 106   | 0.00     |
| 31 T  | Cyclohexane                 | 0.949 | 0.816 | 14.0   | 102   | 0.00     |
| 32 T  | 1,1,1-Trichloroethane       | 0.752 | 0.752 | 0.0    | 104   | 0.00     |
| 33 S  | 1,2-Dichloroethane-d4       | 0.502 | 0.441 | 12.2   | 93    | 0.00     |
| 34 I  | 1,4-Difluorobenzene         | 1.000 | 1.000 | 0.0    | 99    | 0.00     |
| 35 S  | Dibromofluoromethane        | 0.316 | 0.297 | 6.0    | 94    | 0.00     |
| 36 T  | 1,1-Dichloropropene         | 0.451 | 0.483 | -7.1   | 107   | 0.00     |
| 37 T  | Ethyl Acetate               | 0.252 | 0.261 | -3.6   | 96    | 0.00     |
| 38 T  | Carbon Tetrachloride        | 0.454 | 0.473 | -4.2   | 105   | 0.00     |
| 39 T  | Methylcyclohexane           | 0.535 | 0.578 | -8.0   | 105   | 0.00     |
| 40 TM | Benzene                     | 1.390 | 1.433 | -3.1   | 104   | 0.00     |
| 41 T  | Methacrylonitrile           | 0.141 | 0.153 | -8.5   | 109   | 0.00     |
| 42 TM | 1,2-Dichloroethane          | 0.413 | 0.424 | -2.7   | 105   | 0.00     |
| 43 T  | Isopropyl Acetate           | 0.486 | 0.465 | 4.3    | 99    | 0.00     |
| 44 TM | Trichloroethene             | 0.396 | 0.408 | -3.0   | 108   | 0.00     |
| 45 C  | 1,2-Dichloropropane         | 0.356 | 0.376 | -5.6#  | 105   | 0.00     |

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN021119\  
 Data File : VN053694.D  
 Acq On : 11 Feb 2019 9:57  
 Operator : JC/SP  
 Sample : VSTDCCC050  
 Misc : 5.00mL/MSVOA N/WATER  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 LabSampleId :  
 VSTDCCC050

Quant Time: Feb 11 22:22:08 2019  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_N\METHODS\82N013019W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Feb 01 09:11:52 2019  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|       | Compound                    | AvqRF | CCRF  | %Dev  | Area% | Dev(min) |
|-------|-----------------------------|-------|-------|-------|-------|----------|
| 46 T  | Dibromomethane              | 0.214 | 0.217 | -1.4  | 104   | 0.00     |
| 47 T  | Bromodichloromethane        | 0.446 | 0.474 | -6.3  | 108   | 0.00     |
| 48 T  | Methyl methacrylate         | 0.232 | 0.222 | 4.3   | 96    | 0.00     |
| 49 T  | 1,4-Dioxane                 | 0.003 | 0.003 | 0.0   | 106   | 0.00     |
| 50 S  | Toluene-d8                  | 1.186 | 1.108 | 6.6   | 94    | 0.00     |
| 51 T  | 4-Methyl-2-Pentanone        | 0.239 | 0.251 | -5.0  | 101   | 0.00     |
| 52 CM | Toluene                     | 0.825 | 0.869 | -5.3# | 104   | 0.00     |
| 53 T  | t-1,3-Dichloropropene       | 0.446 | 0.471 | -5.6  | 104   | 0.00     |
| 54 T  | cis-1,3-Dichloropropene     | 0.529 | 0.562 | -6.2  | 105   | 0.00     |
| 55 T  | 1,1,2-Trichloroethane       | 0.297 | 0.315 | -6.1  | 105   | 0.00     |
| 56 T  | Ethyl methacrylate          | 0.352 | 0.366 | -4.0  | 99    | 0.00     |
| 57 T  | 1,3-Dichloropropane         | 0.508 | 0.528 | -3.9  | 105   | 0.00     |
| 58 T  | 2-Chloroethyl Vinyl ether   | 0.196 | 0.076 | 61.2# | 36#   | 0.00     |
| 59 T  | 2-Hexanone                  | 0.162 | 0.172 | -6.2  | 104   | 0.00     |
| 60 T  | Dibromochloromethane        | 0.335 | 0.363 | -8.4  | 107   | 0.00     |
| 61 T  | 1,2-Dibromoethane           | 0.294 | 0.296 | -0.7  | 102   | 0.00     |
| 62 S  | 4-Bromofluorobenzene        | 0.396 | 0.378 | 4.5   | 96    | 0.00     |
| 63 I  | Chlorobenzene-d5            | 1.000 | 1.000 | 0.0   | 100   | 0.00     |
| 64 T  | Tetrachloroethene           | 0.429 | 0.438 | -2.1  | 105   | 0.00     |
| 65 PM | Chlorobenzene               | 1.062 | 1.085 | -2.2  | 105   | 0.00     |
| 66 T  | 1,1,1,2-Tetrachloroethane   | 0.371 | 0.394 | -6.2  | 107   | 0.00     |
| 67 C  | Ethyl Benzene               | 1.745 | 1.822 | -4.4# | 105   | 0.00     |
| 68 T  | m/p-Xylenes                 | 0.666 | 0.711 | -6.8  | 106   | 0.00     |
| 69 T  | o-Xylene                    | 0.643 | 0.683 | -6.2  | 105   | 0.00     |
| 70 T  | Styrene                     | 1.043 | 1.135 | -8.8  | 107   | 0.00     |
| 71 P  | Bromoform                   | 0.254 | 0.278 | -9.4  | 107   | 0.00     |
| 72 I  | 1,4-Dichlorobenzene-d4      | 1.000 | 1.000 | 0.0   | 103   | 0.00     |
| 73 T  | Isopropylbenzene            | 3.515 | 3.574 | -1.7  | 106   | 0.00     |
| 74 T  | N-amyl acetate              | 0.845 | 0.828 | 2.0   | 100   | 0.00     |
| 75 P  | 1,1,2,2-Tetrachloroethane   | 0.841 | 0.779 | 7.4   | 100   | 0.00     |
| 76 T  | 1,2,3-Trichloropropane      | 0.642 | 0.653 | -1.7  | 109   | 0.00     |
| 77 T  | Bromobenzene                | 0.947 | 0.939 | 0.8   | 104   | 0.00     |
| 78 T  | n-propylbenzene             | 4.063 | 4.234 | -4.2  | 108   | 0.00     |
| 79 T  | 2-Chlorotoluene             | 2.411 | 2.472 | -2.5  | 108   | 0.00     |
| 80 T  | 1,3,5-Trimethylbenzene      | 2.915 | 3.135 | -7.5  | 109   | 0.00     |
| 81 T  | trans-1,4-Dichloro-2-butene | 0.228 | 0.230 | -0.9  | 102   | 0.00     |
| 82 T  | 4-Chlorotoluene             | 2.474 | 2.533 | -2.4  | 107   | 0.00     |
| 83 T  | tert-Butylbenzene           | 2.576 | 2.644 | -2.6  | 107   | 0.00     |
| 84 T  | 1,2,4-Trimethylbenzene      | 2.970 | 3.175 | -6.9  | 109   | 0.00     |
| 85 T  | sec-Butylbenzene            | 3.506 | 3.725 | -6.2  | 109   | 0.00     |
| 86 T  | p-Isopropyltoluene          | 3.066 | 3.339 | -8.9  | 111   | 0.00     |
| 87 T  | 1,3-Dichlorobenzene         | 1.703 | 1.727 | -1.4  | 108   | 0.00     |
| 88 T  | 1,4-Dichlorobenzene         | 1.685 | 1.701 | -0.9  | 108   | 0.00     |
| 89 T  | n-Butylbenzene              | 2.654 | 2.936 | -10.6 | 113   | 0.00     |

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN021119\  
Data File : VN053694.D  
Acq On : 11 Feb 2019 9:57  
Operator : JC/SP  
Sample : VSTDCCC050  
Misc : 5.00mL/MSVOA N/WATER  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
LabSampleId :  
VSTDCCC050

Quant Time: Feb 11 22:22:08 2019  
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_N\METHODS\82N013019W.M  
Quant Title : SW846 8260  
QLast Update : Fri Feb 01 09:11:52 2019  
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
Max. RRF Dev : 25% Max. Rel. Area : 150%

|      | Compound                    | AvgRF | CCRF  | %Dev  | Area% | Dev(min) |
|------|-----------------------------|-------|-------|-------|-------|----------|
| 90 T | Hexachloroethane            | 0.506 | 0.548 | -8.3  | 115   | 0.00     |
| 91 T | 1,2-Dichlorobenzene         | 1.630 | 1.652 | -1.3  | 106   | 0.00     |
| 92 T | 1,2-Dibromo-3-Chloropropane | 0.120 | 0.119 | 0.8   | 103   | 0.00     |
| 93 T | 1,2,4-Trichlorobenzene      | 0.935 | 1.012 | -8.2  | 105   | 0.00     |
| 94 T | Hexachlorobutadiene         | 0.582 | 0.647 | -11.2 | 116   | 0.00     |
| 95 T | Naphthalene                 | 1.905 | 1.959 | -2.8  | 94    | 0.00     |
| 96 T | 1,2,3-Trichlorobenzene      | 0.868 | 0.920 | -6.0  | 104   | 0.00     |

(#) = Out of Range

SPCC's out = 0 CCC's out = 6