

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021120\
 Data File : VN060054.D
 Acq On : 11 Feb 2020 10:32
 Operator : JC/MD
 Sample : VN0211WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0211WBL01

Quant Time: Feb 12 04:29:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 18 03:09:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	192605	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	300954	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	265696	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	109137	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	110562	51.69	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.38%	
35) Dibromofluoromethane	7.57	113	91117	50.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.54%	
50) Toluene-d8	10.08	98	356056	51.20	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.40%	
62) 4-Bromofluorobenzene	12.40	95	120410	48.03	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.06%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021120\
Data File : VN060054.D
Acq On : 11 Feb 2020 10:32
Operator : JC/MD
Sample : VN0211WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0211WBL01

Quant Time: Feb 12 04:29:36 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
Quant Title : SW846 8260
QLast Update : Sat Jan 18 03:09:32 2020
Response via : Initial Calibration

