


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN021120\
 Data File : VN060052.D
 Acq On : 11 Feb 2020 9:20
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
 Title : SW846 8260
 Last Update : Sat Jan 18 03:09:32 2020



AutoFind: Scans 3303, 3304, 3305; Background Corrected with Scan 3291

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.7	5006	PASS
75	95	30	60	50.3	12747	PASS
95	95	100	100	100.0	25354	PASS
96	95	5	9	7.8	1965	PASS
173	174	0.00	2	0.4	92	PASS
174	95	50	100	84.6	21445	PASS
175	174	5	9	7.6	1628	PASS
176	174	95	101	100.4	21525	PASS
177	176	5	9	6.5	1398	PASS