

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021121\
 Data File : VN065809.D
 Acq On : 11 Feb 2021 12:26
 Operator : JC/MD
 Sample : VN0211MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0211MBL01

Quant Time: Feb 11 13:33:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.621	168	150063	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.547	114	240844	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	210656	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.312	152	72355	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.984	65	80722	40.89	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	81.78%
35) Dibromofluoromethane	7.544	113	69453	48.66	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.32%
50) Toluene-d8	10.055	98	296730	52.89	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.78%
62) 4-Bromofluorobenzene	12.373	95	90511	44.80	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	89.60%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021121\
Data File : VN065809.D
Acq On : 11 Feb 2021 12:26
Operator : JC/MD
Sample : VN0211MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0211MBL01

Quant Time: Feb 11 13:33:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

