

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021121\
 Data File : VN065814.D
 Acq On : 11 Feb 2021 14:48
 Operator : JC/MD
 Sample : M1363-04
 Misc : 2.69g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VCP-SS-10-SOIL

Quant Time: Feb 11 15:10:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.618	168	155184	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.544	114	250681	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	238071	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	100311	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.981	65	85377	41.82	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	83.64%
35) Dibromofluoromethane	7.544	113	71747	48.30	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.60%
50) Toluene-d8	10.055	98	312417	53.50	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.00%
62) 4-Bromofluorobenzene	12.373	95	111702	53.12	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	106.24%
Target Compounds						
					Qvalue	
68) m/p-Xylenes	11.595	106	3603	1.06	ug/l	90
78) n-propylbenzene	12.560	91	3366	0.33	ug/l	98
84) 1,2,4-Trimethylbenzene	13.010	105	4975	0.67	ug/l	96
85) sec-Butylbenzene	13.145	105	5028	0.60	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021121\
Data File : VN065814.D
Acq On : 11 Feb 2021 14:48
Operator : JC/MD
Sample : M1363-04
Misc : 2.69g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VCP-SS-10-SOIL

Quant Time: Feb 11 15:10:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

