

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021121\
 Data File : VN065817.D
 Acq On : 11 Feb 2021 16:03
 Operator : JC/MD
 Sample : M1363-02ME
 Misc : 2.79g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VCP-SS-8-SOILME

Quant Time: Feb 11 16:36:39 2021
 Quant Title :
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.618	168	192287	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.544	114	323961	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	292390	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.312	152	134991	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.984	65	105715	41.79	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	83.58%
35) Dibromofluoromethane	7.544	113	89334	46.53	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.06%
50) Toluene-d8	10.055	98	414975	54.99	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	109.98%
62) 4-Bromofluorobenzene	12.373	95	140057	51.54	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	103.08%
Target Compounds						
					Qvalue	
68) m/p-Xylenes	11.595	106	13869	3.33	ug/l	85
73) Isopropylbenzene	12.219	105	35958	3.04	ug/l	75
78) n-propylbenzene	12.560	91	106102	7.62	ug/l	97
80) 1,3,5-Trimethylbenzene	12.701	105	32051	3.18	ug/l	93
84) 1,2,4-Trimethylbenzene	13.010	105	147241	14.67	ug/l	99
85) sec-Butylbenzene	13.142	105	116753	10.31	ug/l	94
86) p-Isopropyltoluene	13.257	119	34831	3.48	ug/l	92
89) n-Butylbenzene	13.585	91	128570	14.31	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021121\
Data File : VN065817.D
Acq On : 11 Feb 2021 16:03
Operator : JC/MD
Sample : M1363-02ME
Misc : 2.79g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VCP-SS-8-SOILME

Quant Time: Feb 11 16:36:39 2021
Quant Title :
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

