

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021121\
 Data File : VN065821.D
 Acq On : 11 Feb 2021 18:08
 Operator : JC/MD
 Sample : M1388-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB

Quant Time: Feb 12 01:30:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.621	168	208859	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.547	114	357515	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	306990	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.312	152	104395	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.984	65	119564	43.51	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	87.02%
35) Dibromofluoromethane	7.547	113	100874	47.61	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.22%
50) Toluene-d8	10.055	98	442265	53.11	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.22%
62) 4-Bromofluorobenzene	12.370	95	138977	46.34	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	92.68%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021121\
Data File : VN065821.D
Acq On : 11 Feb 2021 18:08
Operator : JC/MD
Sample : M1388-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB

Quant Time: Feb 12 01:30:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

