

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021121\
 Data File : VN065827.D
 Acq On : 11 Feb 2021 20:36
 Operator : JC/MD
 Sample : M1384-04
 Misc : 1.10g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DRUM-2

Quant Time: Feb 12 01:32:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.618	168	200240	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.544	114	344511	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	313167	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	126520	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.984	65	115026	43.66	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	87.32%
35) Dibromofluoromethane	7.547	113	95358	46.71	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.42%
50) Toluene-d8	10.055	98	429421	53.51	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.02%
62) 4-Bromofluorobenzene	12.373	95	150930	52.23	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	104.46%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021121\
Data File : VN065827.D
Acq On : 11 Feb 2021 20:36
Operator : JC/MD
Sample : M1384-04
Misc : 1.10g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DRUM-2

Quant Time: Feb 12 01:32:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

