

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021125\
 Data File : VN085737.D
 Acq On : 11 Feb 2025 11:53
 Operator : JC\MD
 Sample : VN0211WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0211WBL01

Quant Time: Feb 11 23:15:10 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	213433	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	391202	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.859	117	340384	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	133336	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	170771	49.567	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.140%
35) Dibromofluoromethane	8.159	113	142670	52.569	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.140%
50) Toluene-d8	10.565	98	480003	49.779	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.560%
62) 4-Bromofluorobenzene	12.847	95	148954	45.158	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	90.320%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021125\
Data File : VN085737.D
Acq On : 11 Feb 2025 11:53
Operator : JC\MD
Sample : VN0211WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0211WBL01

Quant Time: Feb 11 23:15:10 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

