

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021125\
 Data File : VN085747.D
 Acq On : 11 Feb 2025 16:02
 Operator : JC\MD
 Sample : PB166652ZHE#07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB166652ZHE#07

Quant Time: Feb 11 23:21:02 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	249253	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	492314	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	459304	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	170760	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	202163	50.246	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	100.500%	
35) Dibromofluoromethane	8.165	113	168674	49.386	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	98.780%	
50) Toluene-d8	10.565	98	622056	51.261	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	102.520%	
62) 4-Bromofluorobenzene	12.847	95	189006	45.532	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	91.060%	
Target Compounds						
						Qvalue
16) Acetone	4.430	43	39890	30.684	ug/l	90
20) Methylene Chloride	5.277	84	4796	1.490	ug/l #	85
43) Isopropyl Acetate	8.682	43	224070	28.902	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021125\
Data File : VN085747.D
Acq On : 11 Feb 2025 16:02
Operator : JC\MD
Sample : PB166652ZHE#07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB166652ZHE#07

Quant Time: Feb 11 23:21:02 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

