

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021125\
 Data File : VN085754.D
 Acq On : 11 Feb 2025 18:49
 Operator : JC\MD
 Sample : Q1344-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SOIL-1

Quant Time: Feb 11 23:24:25 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	225336	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	448978	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	422884	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	157190	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	186505	51.274	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	102.540%	
35) Dibromofluoromethane	8.165	113	153667	49.334	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	98.660%	
50) Toluene-d8	10.565	98	574156	51.881	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	103.760%	
62) 4-Bromofluorobenzene	12.847	95	171279	45.244	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	90.480%	
Target Compounds						
						Qvalue
16) Acetone	4.430	43	40692	34.623	ug/l	93
20) Methylene Chloride	5.277	84	3442	1.183	ug/l #	71
43) Isopropyl Acetate	8.682	43	227066	32.115	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021125\
Data File : VN085754.D
Acq On : 11 Feb 2025 18:49
Operator : JC\MD
Sample : Q1344-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SOIL-1

Quant Time: Feb 11 23:24:25 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

