

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021125\
 Data File : VN085755.D
 Acq On : 11 Feb 2025 19:13
 Operator : JC\MD
 Sample : Q1344-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SOIL-2

Quant Time: Feb 11 23:24:54 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	214906	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	426116	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	400946	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	144536	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	178658	51.501	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 103.000%		
35) Dibromofluoromethane	8.165	113	150418	50.882	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 101.760%		
50) Toluene-d8	10.565	98	550402	52.403	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 104.800%		
62) 4-Bromofluorobenzene	12.847	95	159357	44.353	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 88.700%		
Target Compounds						
						Qvalue
16) Acetone	4.430	43	41164	36.725	ug/l	98
20) Methylene Chloride	5.277	84	4481	1.615	ug/l #	77
43) Isopropyl Acetate	8.682	43	231830	34.548	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021125\
Data File : VN085755.D
Acq On : 11 Feb 2025 19:13
Operator : JC\MD
Sample : Q1344-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SOIL-2

Quant Time: Feb 11 23:24:54 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

