

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021125\
 Data File : VN085760.D
 Acq On : 11 Feb 2025 21:13
 Operator : JC\MD
 Sample : Q1349-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TWP-1-PERMIT

Quant Time: Feb 11 23:27:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	212008	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	414040	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	359182	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	145329	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	179528	52.459	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery	= 104.920%		
35) Dibromofluoromethane	8.165	113	149931	52.197	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery	= 104.400%		
50) Toluene-d8	10.565	98	499933	48.986	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery	= 97.980%		
62) 4-Bromofluorobenzene	12.847	95	155780	44.622	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery	= 89.240%		
Target Compounds						Qvalue
16) Acetone	4.430	43	5390	4.874	ug/l	98
19) Methyl tert-butyl Ether	5.800	73	4971	0.673	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021125\
Data File : VN085760.D
Acq On : 11 Feb 2025 21:13
Operator : JC\MD
Sample : Q1349-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TWP-1-PERMIT

Quant Time: Feb 11 23:27:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

