

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021219\
 Data File : VN053724.D
 Acq On : 12 Feb 2019 12:25
 Operator : JC/SP
 Sample : PB117023TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB117023TB

Quant Time: Feb 13 04:43:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	713038	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1162611	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1071076	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	422102	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	376419	52.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.24%	
35) Dibromofluoromethane	7.59	113	368734	50.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.38%	
50) Toluene-d8	10.09	98	1412560	51.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.42%	
62) 4-Bromofluorobenzene	12.40	95	465373	50.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.18%	
Target Compounds						
16) Acetone	3.82	43	26472	14.428	ug/l	99
20) Methylene Chloride	4.55	84	18453	2.397	ug/l	90
43) Isopropyl Acetate	8.17	43	33536	2.967	ug/l #	88
95) Naphthalene	15.13	128	6532	2.076	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021219\
Data File : VN053724.D
Acq On : 12 Feb 2019 12:25
Operator : JC/SP
Sample : PB117023TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB117023TB

Quant Time: Feb 13 04:43:44 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

