

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021219\
 Data File : VN053733.D
 Acq On : 12 Feb 2019 16:27
 Operator : JC/SP
 Sample : K1459-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 M-10-D

Manual Integrations
 APPROVED

MMDadoda
 2/13/2019 12:15:22 PM

Quant Time: Feb 13 05:50:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	729554	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1209342	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1116130	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	459070	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	387989	53.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.02%	
35) Dibromofluoromethane	7.59	113	382697	50.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.16%	
50) Toluene-d8	10.09	98	1466599	51.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.22%	
62) 4-Bromofluorobenzene	12.40	95	511776	53.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.96%	
Target Compounds						
16) Acetone	3.82	43	13047	6.950	ug/l	90
20) Methylene Chloride	4.55	84	700004	88.878	ug/l	98
43) Isopropyl Acetate	8.18	43	28562m	2.429	ug/l	
95) Naphthalene	15.13	128	18024	2.599	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021219\
Data File : VN053733.D
Acq On : 12 Feb 2019 16:27
Operator : JC/SP
Sample : K1459-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
M-10-D

Manual Integrations
APPROVED

MMDadoda
2/13/2019 12:15:22 PM

Quant Time: Feb 13 05:50:15 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

