

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021219\
 Data File : VN053734.D
 Acq On : 12 Feb 2019 16:54
 Operator : JC/SP
 Sample : K1459-12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 M-11-S

Quant Time: Feb 13 05:51:49 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	731873	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1209383	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1112515	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	461184	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	393081	53.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.08%	
35) Dibromofluoromethane	7.59	113	383199	50.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.30%	
50) Toluene-d8	10.09	98	1453944	50.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.34%	
62) 4-Bromofluorobenzene	12.40	95	505174	52.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.58%	
Target Compounds						
16) Acetone	3.82	43	19180	10.184	ug/l	97
20) Methylene Chloride	4.55	84	221052	27.977	ug/l	96
43) Isopropyl Acetate	8.17	43	45963	3.909	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021219\
Data File : VN053734.D
Acq On : 12 Feb 2019 16:54
Operator : JC/SP
Sample : K1459-12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
M-11-S

Quant Time: Feb 13 05:51:49 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

