

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021219\
 Data File : VN053736.D
 Acq On : 12 Feb 2019 17:47
 Operator : JC/SP
 Sample : K1461-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TEST-PIT-4

Manual Integrations
 APPROVED

MMDadoda
 2/13/2019 12:15:30 PM

Quant Time: Feb 13 05:55:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	771321	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1258284	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1174856	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	528522	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	400216	51.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.44%	
35) Dibromofluoromethane	7.59	113	392734	49.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.80%	
50) Toluene-d8	10.09	98	1518557	50.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.74%	
62) 4-Bromofluorobenzene	12.40	95	516219	51.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.70%	
Target Compounds						
16) Acetone	3.82	43	42372	21.349	ug/l	97
20) Methylene Chloride	4.55	84	1144649	137.463	ug/l	98
25) 2-Butanone	6.84	43	7103	2.509	ug/l #	84
43) Isopropyl Acetate	8.19	43	48292m	3.947	ug/l	
95) Naphthalene	15.13	128	125670	6.961	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021219\
Data File : VN053736.D
Acq On : 12 Feb 2019 17:47
Operator : JC/SP
Sample : K1461-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TEST-PIT-4

Manual Integrations
APPROVED

MMDadoda
2/13/2019 12:15:30 PM

Quant Time: Feb 13 05:55:43 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

