

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021219\
 Data File : VN053740.D
 Acq On : 12 Feb 2019 19:34
 Operator : JC/SP
 Sample : K1462-16
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-2-D

Quant Time: Feb 13 06:05:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	734055	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1205988	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1184629	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	507662	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	392908	53.36	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.72%	
35) Dibromofluoromethane	7.59	113	385992	50.65	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.30%	
50) Toluene-d8	10.09	98	1470794	51.40	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.80%	
62) 4-Bromofluorobenzene	12.40	95	526433	55.17	ug/l	0.00
Spiked Amount	50.000		Recovery	=	110.34%	
Target Compounds						
16) Acetone	3.83	43	10579	5.601	ug/l	100
20) Methylene Chloride	4.56	84	556147	70.180	ug/l	98
43) Isopropyl Acetate	8.17	43	40027	3.414	ug/l #	74
68) m/p-Xylenes	11.62	106	26491	1.679	ug/l	96
80) 1,3,5-Trimethylbenzene	12.73	105	31346	1.059	ug/l	100
84) 1,2,4-Trimethylbenzene	13.04	105	114828	3.808	ug/l	98
95) Naphthalene	15.13	128	152037	8.317	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021219\
Data File : VN053740.D
Acq On : 12 Feb 2019 19:34
Operator : JC/SP
Sample : K1462-16
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-2-D

Quant Time: Feb 13 06:05:01 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

