

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021219\
 Data File : VN053741.D
 Acq On : 12 Feb 2019 20:01
 Operator : JC/SP
 Sample : K1462-20
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-3-S

Quant Time: Feb 13 06:06:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	702182	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1175766	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1079916	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	440938	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	373954	53.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.18%	
35) Dibromofluoromethane	7.59	113	374153	50.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.72%	
50) Toluene-d8	10.09	98	1414372	50.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.40%	
62) 4-Bromofluorobenzene	12.40	95	468508	50.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.72%	
Target Compounds						
16) Acetone	3.82	43	11475	6.351	ug/l	95
20) Methylene Chloride	4.55	84	2240114	295.509	ug/l	99
43) Isopropyl Acetate	8.18	43	33467	2.928	ug/l #	80
95) Naphthalene	15.14	128	6920	2.081	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021219\
Data File : VN053741.D
Acq On : 12 Feb 2019 20:01
Operator : JC/SP
Sample : K1462-20
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-3-S

Quant Time: Feb 13 06:06:19 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

