

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021219\
 Data File : VN053742.D
 Acq On : 12 Feb 2019 20:28
 Operator : JC/SP
 Sample : K1462-24
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-3-D

Manual Integrations
 APPROVED

MMDadoda
 2/13/2019 12:15:41 PM

Quant Time: Feb 13 06:08:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	729957	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1199356	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1094487	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	440979	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	381934	52.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.32%	
35) Dibromofluoromethane	7.59	113	378676	49.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.94%	
50) Toluene-d8	10.09	98	1451104	51.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.00%	
62) 4-Bromofluorobenzene	12.40	95	478536	50.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.84%	
Target Compounds						
16) Acetone	3.82	43	12045m	6.413	ug/l	Qvalue
20) Methylene Chloride	4.55	84	2515083	319.157	ug/l	99
43) Isopropyl Acetate	8.19	43	48746m	4.180	ug/l	
95) Naphthalene	15.13	128	23502	2.907	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021219\
Data File : VN053742.D
Acq On : 12 Feb 2019 20:28
Operator : JC/SP
Sample : K1462-24
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-3-D

Manual Integrations
APPROVED

MMDadoda
2/13/2019 12:15:41 PM

Quant Time: Feb 13 06:08:07 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

